



Product Service

C E R T I F I C A T E

No. Z1A 17 12 12906 250

Holder of Certificate: DÜPERTHAL SICHERHEITSTECHNIK
GMBH & Co. KGFrankenstraße 3
63791 Karlstein
GERMANY**DÜPERTHAL®****Factory(ies):**

62099

Certification Mark:**Product:****Safety cabinets****Model(s):****COOL line 89-20097u-xyz(ss)****Parameters:**

Key code Art. No.

u: Door technology, stands for the numbers 0, 7, 9
x: colour, stands for the numbers 0, 4, 5, 6, 7
yz: interior fitting, stands for the numbers 00-99
ss: special sign, stands for B, L

fire resistance: FWF 90

The article-no. exactly defines material and size
of the single products. There is a detailed description
of the characteristics in the technical report.

Tested according to:

DIN EN 14470-1:2004
DIN EN 14727:2006
TRGS 510:2015 Anlage 3
EK5/AK4 09-10:2009
AfPS GS 2014:01 PAK

The product meets the safety and health requirements of the German Product Safety Act section 20 to 22 ProdSG. The certification marks shown above can be affixed on the product. It is not permitted to alter the certification marks in any way. In addition the certificate holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. See also notes overleaf.

Test report no.:

713111034

Valid until:

2022-12-18

Date, 2017-12-19

(Horst Kristen)

Page 1 of 1

