



Product Service

CERTIFICATE

No. Z1A 012906 0508 Rev. 00

Holder of Certificate:



**DÜPERTHAL SICHERHEITSTECHNIK
GMBH & Co. KG**

Frankenstraße 3
63791 Karlstein
GERMANY

Factories:

062099

Certification Mark:



Product:

Safety cabinets

Model(s):

CHEM standard, CHEM tox

Parameters:

CHEM standard CHEM tox	
model type	approx. size in mm (width x height x depth)
M	594 x 2045 x 612
L	894 x 2045 x 612
XL	1194 x 2045 x 612

Class of fire resistance: FWF 0

A detailed description of the specifications can
be found in the test report.

Tested according to:

PPP 52125B:2019
DIN EN 16121:2017
DIN EN 16122:2012

The product meets the safety and health requirements of the German Product Safety Act section 20 to 22 ProdSG. The certification marks shown above can be affixed on the product. It is not permitted to alter the certification marks in any way. In addition, the certificate holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the testing and certification regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.:

713144996

Valid until:

2025-04-16

Date,

2020-06-02

(Frank Feihle)